

## The Standards Institution of Israel

This is to certify that:

ElNet Energy & Power Quality Powermeter

Models:

ElNet PQ-GR, ElNet LTC, ElNet PFC,

ElNet LT/LTE, ElNet PICO, ElNet MC,

ElNet VIP

Ratings:

100-240Vac, 50/60Hz, 20VA

Measurement circuits: Voltage: 0-650Vac.

Max. input voltage: 1000V,

Max. input current: 6A

Manufactured by: Control Applications Ltd.

Address:

2A Habarzel St., Tel Aviv,

0069710, Israel

have been tested by the Standards Institution of Israel and found to comply with the standard requirements of:

❖IEC 61010-1:2010 / EN 61010-1: 2010

Safety requirements for electrical equipment for measurement, control, and laboratory use. Part 1: General requirements

(Test results are detailed in SII Test Report No.: 9312314380.)

CONFORMITY

Certificate No.: 9312314380

Michael Terman

Head of Electrical Safety Branch

Date of issue: 07/05/2015

Electronics and Telematics Laboratory

42 Chaim Levanon st. Tel-Aviv 69977, Israel. Telematics Laboratory. Tel. 972-3-6467800 Fax. 972-3-6467779